

(FOR RESEARCH USE ONLY. DO NOT USE IT IN CLINICAL DIAGNOSIS!)

SARS-COV-2 Spike Monoclonal Antibody(IgG / IgM Positive Control)

Catalog No: E-AB-V1007

Size: $200 \, \mu L / 500 \, \mu L / 1 \, mL$

Cat.	Products	200 μL	500 μL	1 mL	Storage
E-AB-V1007A	SARS-COV-2 Spike Monoclonal	200 µL	500 μL	1 mL	-20°C
	Antibody(Human IgG, #5D9)				
E-AB-V1007B	SARS-COV-2 Spike Monoclonal	200 µL	500 μL	1 mL	-20°C
	Antibody(Human IgM, #9E8)				
Manual		One Copy			

This manual must be read attentively and completely before using this product.

If you have any problems, please contact our Technical Service Center for help.

Phone: 240-252-7368(USA) Fax: 240-252-7376(USA)

Email: techsupport@elabscience.com

Website: www.elabscience.com

Please kindly provide us the lot number (on the outside of the box) of the kit for more efficient service.

Introduction

The product contains two isotype antibody clone Human IgG and IgM, which can recognizes 2019-nCoV S-RBD Protein.

Source

Cultured in vitro under conditions free from animal derived components.

Recommended diluent

- 1. 50 fold diluted healthy human control plasma or serum with PBS (pH 7.4).
- 2. PBS (pH 7.4), 4% BSA, 0.2% proclin300 is recommended as preservative.


Intented use


The product is intended for the quality control of serum on clinical chemistry and immunoassay systems. The assayed liquid protein controls are for the control of accuracy. For TRFIA assay, antibody can be diluted to 1:50~1:100, the dilution rate should be adjusted based on different testing method.

Storage

Store at -20°C . Avoid of freezing. Valid for 12 months.

Elisa Assay (Nucleocapsid Protein binding)





TRFIA assay

