

(FOR RESEARCH USE ONLY. DO NOT USE IT IN CLINICAL DIAGNOSIS !)

Elabsience[®]Cell Total Iron Colorimetric Assay Kit

Catalog No: E-BC-K880-M

Specification: 48T(32 samples)/96T(80 samples)

Measuring instrument: Microplate reader (590-600 nm)

Detection range: 0.4 -50 $\mu\text{mol/L}$

This manual must be read attentively and completely before using this product.

If you have any problem, please contact our Technical Service Center for help:

Toll-free: 1-888-852-8623

Tell: 1-832-243-6086

Fax: 1-832-243-6017

Email: techsupport@elabsience.com


Website: www.elabsience.com

Please kindly provide us the lot number (on the outside of the box) of the kit for more efficient service.

Table of contents

Assay summary	3
Intended use	4
Detection principle	4
Kit components & storage	4
Materials prepared by users	5
Reagent preparation	5
Sample preparation	6
The key points of the assay	6
Operating steps	7
Calculation	7
Appendix I Performance Characteristics	8
Appendix II Example Analysis	10
Statement	11

Assay summary



Intended use

This kit can measure total iron content in cell sample.

Detection principle

Iron is one of the metal elements in organism and has important physiological functions. Ferrous ion is a key element in heme and hemoglobin and plays an important role in many biochemical reactions. Under the action of reductant, iron ions in samples can be reduced into ferrous ions (Fe^{2+}). The latter then bind to probe and form complexes, which has a maximum absorption peak at 593 nm. The concentration of iron can be calculated by measuring the OD value at 593 nm indirectly.

Kit components & storage

Item	Component	Size 1(48 T)	Size 2(96 T)	Storage
Reagent 1	Buffer Solution	25 mL × 1 vial	50 mL × 1 vial	2-8°C, 12 month shading light
Reagent 2	Chromogenic Solution	5 mL × 1 vial	10 mL × 1 vial	2-8°C, 12 month shading light s
Reagent 3	10 mmol/L Iron Standard	1 mL × 1 vial	1 mL × 2 vials	2-8°C, 12 month shading light
	Microplate	96 wells		No requirement
	Plate Sealer	2 pieces		

Note: The reagents must be stored strictly according to the preservation conditions in the above table. The reagents in different kits cannot be mixed with each other. For a small volume of reagents, please centrifuge before use, so as not to obtain sufficient amount of reagents.

Materials prepared by users

Instruments:

Centrifuge, Incubator, Microplate reader (590-600 nm, optimum wavelength: 593 nm)

Reagents:

Double distilled water

Reagent preparation

① Equilibrate all reagents to room temperature before use.

② Preparation of 100 $\mu\text{mol/L}$ iron standard:

Dissolve 20 μL of 10 mmol/L iron standard with 1980 μL of double distilled water, mix well to dissolve. The 100 $\mu\text{mol/L}$ iron standard should be prepared on spot.

③ The preparation of standard curve:

Always prepare a fresh set of standards. Discard working standard dilutions after use.

Dilute 100 $\mu\text{mol/L}$ iron standard with buffer solution to a serial concentration. The recommended dilution gradient is as follows: 0, 5, 10, 15, 20, 30, 40, 50 $\mu\text{mol/L}$. Reference is as follows:

Item	①	②	③	④	⑤	⑥	⑦	⑧
Concentration ($\mu\text{mol/L}$)	0	5	10	15	20	30	40	50
100 $\mu\text{mol/L}$ iron standard (μL)	0	50	100	150	200	300	400	500
Buffer solution (μL)	1000	950	900	850	800	80	600	500

Sample preparation

① Sample preparation

Cell (adherent or suspension) samples:

- ① Harvest the number of cells needed for each assay (initial recommendation 1×10^6 cells).
- ② Wash cells with PBS (0.01 M, pH 7.4).
- ③ Lyse 1×10^6 cells with 200 μ L buffer solution. Place on the ice box and mix well every 5 min, lyse for 10 min.
- ④ Centrifuge at $15000 \times g$ for 10 min. Collect supernatant and keep it on ice for detection.

② Dilution of sample

The recommended dilution factor for different samples is as follows (for reference only):

Sample type	Dilution factor
HepG2 Cell	1
molt-4 Cell	1
Jurkat Cell	1
HEL Cell	1

Note: The diluent is buffer solution. For the dilution of other sample types, please do pretest to confirm the dilution factor

The key points of the assay

- ① To avoid contamination, it is recommended to aliquot the chromogenic solution into smaller quantities before use.
- ② Prevent the formulation of bubbles when the reagent or sample is transferred into the microplate.
- ③ Select fresh cell samples for detection.

Operating steps

- ① Standard well: Take 80 μL of standard solution with different concentrations to the corresponding wells.
Sample well: Take 80 μL of sample to the corresponding wells.
- ③ Add 80 μL of chromogenic solution to each well.
- ④ Mix fully and incubate at 37°C for 40 min.
- ⑤ Measure the OD value of each well with microplate reader at 593 nm..

Calculation

The standard curve:

1. Average the duplicate reading for each standard.
2. Subtract the mean OD value of the blank (Standard #①) from all standard readings. This is the absolute OD value.
3. Plot the standard curve by using absolute OD value of standard and correspondent concentration as y-axis and x-axis respectively. Create the standard curve ($y = ax + b$) with graph software (or EXCEL).

The sample:

Cell sample:

$$\text{Fe content (nmol/10}^6) = \frac{\Delta A - b}{a} \div \frac{N}{V} \times f$$

[Note]

ΔA : $\text{OD}_{\text{Sample}} - \text{OD}_{\text{Blank}}$ (OD_{Blank} is the OD value when the standard concentration is 0).

N: The number of cell sample/ 10^6 .

V: The volume of buffer solution in the preparation step of cell, mL.

f: Dilution factor of sample before test.

Appendix I Performance Characteristics

1. Parameter:

Intra-assay Precision

Three HepG2 cell samples were assayed in replicates of 20 to determine precision within an assay (CV = Coefficient of Variation).

Parameters	Sample 1	Sample 2	Sample 3
Mean ($\mu\text{mol/L}$)	3.60	18.70	42.50
%CV	1.6	1.3	1.0

Inter-assay Precision

Three HepG2 cell samples were assayed 20 times in duplicate by three operators to determine precision between assays.

Parameters	Sample 1	Sample 2	Sample 3
Mean ($\mu\text{mol/L}$)	3.60	18.70	42.50
%CV	1.2	1.6	1.7

Recovery

Take three samples of high concentration, middle concentration and low concentration to test the samples of each concentration for 6 times parallelly to get the average recovery rate of 99%.

	Standard 1	Standard 2	Standard 3
Expected Conc. ($\mu\text{mol/L}$)	7.5	18	35
Observed Conc. ($\mu\text{mol/L}$)	7.6	17.6	34.3
Recovery rate (%)	101	98	98


Sensitivity

The analytical sensitivity of the assay is $0.4 \mu\text{mol/L}$. This was determined by adding two standard deviations to the mean O.D. obtained when the zero standard was assayed 20 times, and calculating the corresponding concentration.

2. Standard curve:

As the OD value of the standard curve may vary according to the conditions of the actual assay performance (e.g. operator, pipetting technique or temperature effects), so the standard curve and data are provided as below for reference only:

Concentration ($\mu\text{mol/L}$)	0	5	10	15	20	30	40	50
Average OD	0.050	0.086	0.119	0.154	0.188	0.259	0.325	0.402
Absoluted OD	0.000	0.036	0.070	0.104	0.138	0.210	0.275	0.353



Appendix II Example Analysis


Example analysis:

For HepG2 cell, take 80 μL of the supernatant, and carry the assay according to the operation table. The results are as follows:

Standard curve: $y = 0.00699x - 0.0004$, the average OD value of the sample is 0.070, the average OD value of the blank is 0.050, and the calculation result is:

$$\text{Fe content (nmol}/10^6) = (0.070 - 0.050 + 0.0004) \div 0.00699 \div 1.5 \times 0.2 = 0.39 \text{ nmol}/10^6$$

Detect HepG2 cell, Jurkat cell, Jurkat cell, Molt-4 and 293T cells according to the protocol, the result is as follows:



Statement

1. This assay kit is for Research Use Only. We will not response for any arising problems or legal responsibilities causing by using the kit for clinical diagnosis or other purpose.
2. Please read the instructions carefully and adjust the instruments before the experiments. Please follow the instructions strictly during the experiments.
3. Protection methods must be taken by wearing lab coat and latex gloves.
4. If the concentration of substance is not within the detection range exactly, an extra dilution or concentration should be taken for the sample.
5. It is recommended to take a pre-test if your sample is not listed in the instruction book.
6. The experimental results are closely related to the situation of reagents, operations, environment and so on. Elabscience will guarantee the quality of the kits only, and NOT be responsible for the sample consumption caused by using the assay kits. It is better to calculate the possible usage of sample and reserve sufficient samples before use.

