Elabscience®

6th Edition, revised in Jan, 2021

((FOR RESEARCH USE ONLY. DO NOT USE IT IN CLINICAL DIAGNOSTICS!)

10 × EDTA Anticoagulant

Catalog No: E-EL-SR003 Product size: 30mL

If you have any problems, please contact our Technical Service Center for help (info in the header of each page).

Tel: 1-832-243-6086

Fax: 1-832-243-6017

Email: techsupport@elabscience.com

Website: www.elabscience.com

Use

The reagent is used for anticoagulation during plasma sample collection in ELISA experiments.

Storage

The reagent can be stably stored at room temperature for 6 months.

Background

EDTA-Na₂ binds to calcium ions in the blood to form a chelate that prevents blood from clotting.

Method of use

The effective volume ratio of $10 \times EDTA$ -Na₂ to whole blood is 1:9. For example, 0.2mL EDTA-Na₂ anticoagulant ($10 \times$) can anticoagulate 1.8 mL whole blood.

Plasma sample preparation method

After the whole blood was collected, an appropriate amount of 10×EDTA anticoagulant was added immediately for anticoagulation (if the collection vessel containing anticoagulant was used, this operation would be ignored), the sample was centrifuged at 1000×g for 15 minutes within 30 minutes after collection, and collect the supernatant to carry out the assay. Avoid using hemolysis, hyperlipidemia samples.

Notes on samples

- 1. Blood collection tubes should be disposable and be no-endotoxin.
- 2. Samples should be assayed within 7 days when stored at $2-8^{\circ}$ C, otherwise samples must be divided up and stored at -20° C (≤ 1 month) or -80° C (≤ 3 months). Avoid repeated freeze-thaw cycles.

Note

- 1. This product is proved to be used for ELISA sample collection, and it is recommended to use it with ELISA kit products provided by our company.
- 2. Ensure aseptic operation during operation.
- 3. The product is designed for research use only, not for clinical diagnosis and treatment.