

## (FOR RESEARCH USE ONLY, DO NOT USE IT IN CLINICAL DIAGNOSTICS)

## **Ancillary reagents for ELISA Kits**

This manual must be read thoroughly before using this product.

If you have any problems, please contact our Technical Service Center for help (information in the header of each page).

Phone: 240-252-7368(USA) 240-252-7376(USA)

Email: techsupport@elabscience.com

Website: www.elabscience.com

Please refer to specific expiry date from label on the bottle.

Kindly provide us with the lot number (on the label) of the product for more efficient service.

## Version 1, revised at Jun, 2020

| Catalog No./Product name | E-ELIR-001 ELISA Plate Coating Buffer(1×)                               |
|--------------------------|---|
| Specifications           | 100mL, 1000mL   |
| Description              | Ready to use. For ELISA plate coating with antigen or antibody, can     |
|                          | reduce the coating concentration for antigen or antibody significantly. |
| Method for use           | Use this product to dilute the antigen or antibody to an appropriate    |
|                          | concentration for coating.  |
|                          | Dosage of coating buffer can be determined by end users depending       |
|                          | on their own coating process.   |
| Storage conditions       | 2-8°C, 12 months  |

| Catalog No./Product name | E-ELIR-002 ELISA Plate Coating Buffer(5×)                              |
|--------------------------|--|
| Specifications           | 100mL, 1000mL  |
| Description              | 5×coating buffer. For ELISA plate coating with antigen or antibody,    |
|                          | can reduce the coating concentration for antigen or antibody           |
|                          | significantly.   |
| Method for use           | Prepare the 1× working coating buffer with double distilled water. Use |
|                          | the working buffer to dilute the antigen or antibody to an appropriate |
|                          | concentration for coating.   |
|                          | Dosage of coating buffer can be determined by end users depending      |
|                          | on their own coating process.  |
| Storage conditions       | 2-8°C, 12 months   |

| Catalog No./Product name | E-ELIR-003 ELISA Plate Blocking Buffer                              |
|--------------------------|---|
| Specifications           | 100mL, 1000mL   |
| Description              | Ready to use. For ELISA plate blocking after finishing the coating  |
|                          | process.  |
| Method for use           | 1. After ELISA plate coating process, add 150-300μL blocking        |
|                          | buffer to each well for blocking.                                   |
|                          | 2. The blocked plate can be used for further applications.          |
|                          | 3. For future use, dry the blocked plate and put them to a bag with |
|                          | desiccant and seal up the bag, keep at 2-8℃.                        |
| Storage conditions       | 2-8°C, 12 months  |

| Catalog No./Product name | E-ELIR-004 Wash Buffer for Sandwich-ELISA(25×)                        |
|--------------------------|---|
| Specifications           | 100mL, 1000mL   |
| Description              | 25 × wash buffer. For washing during ELISA assay.                     |
|                          | This wash buffer is suitable for sandwich-ELISA.                      |
| Method for use           | Prepare the 1× working wash buffer with deionized water before use,   |
|                          | discard the solution from each micro plate well, then add about 350µL |

|                    | working wash buffer to the wells, soak for 1-2 minutes, aspirate or decant the solution from each well and pat dry against clean absorbent paper.                      |
|--------------------|--|
| Storage conditions | 2-8°C, 12 months  Note: If there is crystal formation in the concentrate, warm it in a 40°C water bath and mix it gently until the crystals have completely dissolved. |

| Catalog No./Product name | E-ELIR-005 Wash Buffer for Competitive-ELISA(25×)                         |
|--------------------------|---|
| Specifications           | 100mL, 1000mL   |
| Description              | 25× wash buffer. For washing during ELISA assay.                          |
|                          | This wash buffer is suitable for competitive-ELISA.                       |
| Method for use           | Prepare the 1× working wash buffer with deionized water before use,       |
|                          | discard the solution from each micro plate well, then add about 350µL     |
|                          | working wash buffer to the wells, soak for 1-2 minutes, aspirate or       |
|                          | decant the solution from each well and pat dry against clean absorbent    |
|                          | paper.  |
| Storage conditions       | 2-8°C, 12 months  |
|                          | Note: If there is crystal formation in the concentrate, warm it in a 40°C |
|                          | water bath and mix it gently until the crystals have completely           |
|                          | dissolved.  |

| Catalog No./Product name | E-ELIR-006 Stop Solution  |
|--------------------------|---|
| Specifications           | 100mL, 1000mL   |
| Description              | Ready to use. To stop the color reaction in ELISA assay.              |
|                          | The main component is 1M sulfuric acid.                               |
| Method for use           | Add about 50μL to each well for use.                                  |
| Storage conditions       | Room temperature, 12 months   |
|                          | Note: Wear suitable equipment to protect from contacting this reagent |
|                          | directly.   |

| Catalog No./Product name | E-ELIR-007 HRP-conjugate Stabilizer                                   |
|--------------------------|---|
| Specifications           | 10mL, 100mL, 500mL  |
| Description              | Ready to use. Improve the stability of HRP conjugate effectively.     |
| Method for use           | Add this reagent to HRP conjugate and make it over 90% of the         |
|                          | mixture, then it can be stable at 2-8°C for 6 months and at -20°C for |
|                          | 12 months.  |
| Storage conditions       | 2-8°C, 12 months  |

| Catalog No./Product name | E-ELIR-008 HRP-conjugate Diluent   |
|--------------------------|--|
| Specifications           | 100mL, 1000mL  |
| Description              | Ready to use. Use this diluent for HRP conjugate working solution preparation.                             |
| Method for use           | Dilute the concentrated HRP conjugate to 1×HRP conjugate working solution by this product for ELISA assay. |
| Storage conditions       | 2-8°C, 12 months   |

| Catalog No./Product name | E-ELIR-009 Biotinylated Antibody Stabilizer                           |
|--------------------------|---|
| Specifications           | 10mL, 100mL, 500mL  |
| Description              | Ready to use. Improve the stability of biotinylated antibody          |
|                          | effectively.  |
| Method for use           | Add this reagent to biotinylated antibody and make it over 90% of the |
|                          | mixture, then it can be stable at 2-8°C for 6 months and at -20°C for |
|                          | 12 months.  |
| Storage conditions       | 2-8°C, 12 months  |

| Catalog No./Product name | E-ELIR-010 Biotinylated Antibody Diluent  |
|--------------------------|---|
| Specifications           | 100mL, 1000mL   |
| Description              | Ready to use. Use this diluent for biotinylated antibody working solution preparation.                                      |
| Method for use           | Dilute the concentrated biotinylated antibody to 1 xbiotinylated antibody working solution by this product for ELISA assay. |
| Storage conditions       | 2-8°C, 12 months  |

| Catalog No./Product name | E-ELIR-011 Sample Diluent   |
|--------------------------|---|
| Specifications           | 100mL, 1000mL   |
| Description              | Ready to use. Use this diluent for sample dilutions, can reduce the |
|                          | matrix effect significantly.  |
| Method for use           | Dilute tested samples to an appropriate dilution for ELISA assay.   |
|                          | Suitable sample types: serum, plasma, urine, saliva, tissue         |
|                          | homogenate, cell lysate, cell culture supernatant etc.              |
| Storage conditions       | 2-8°C, 12 months  |