

Tel:240-252-7368(USA) Fax: 240-252-7376(USA) techsupport@elabscience.com Website: www.elabscience.com

Elab Fluor® 488 Anti-Human CD20 Antibody[2H7]

Catalog No. E-AB-F1212L Reactivity Human Storage **Applications FCM** Store at 2~8°C, Avoid freeze / thaw cycles

Important Note: Centrifuge before opening to ensure complete recovery of vial contents.

Antigen Information

Alternate Names Bp35,Leukocyte surface antigen Leu-16,MS4A1,B-lymphocyte surface antigen B1

Uniprot ID

Background CD20 is a 33-37 kD, four transmembrane spanning protein, also known as B1 and Bp35. CD20 is

expressed on pre-B-cells, resting and activated B cells (not plasma cells), some follicular dendritic cells, and at low levels on a T cell subset. CD20 is heavily phosphorylated on activated B cells and malignant B cells. Homo-oligomeric complexes of CD20 are thought to form Ca2+ conductive ion channels in the plasma membrane of B cells. The CD20 molecule is involved in B-cell activation and is associated with various Src family kinases (Lyn, Lck, Fyn). It exists in a complex

with MHC class I and II, CD53, CD81, and CD82.

Product Details

Form Liquid

Size 20Tests/100Tests/100Tests×2

Clone No. 2H7 Host Mouse

Isotype Mouse IgG2b, κ

Reactivity Human **FCM Application**

Elab Fluor[®] 488 Mouse IgG2b, κ Isotype Control[MPC-11] [Product E-AB-F09812L] **Isotype Control** Phosphate buffered solution, pH 7.2, containing 0.09% stabilizer and 1% protein protectant. **Storage Buffer**

Shipping Biological ice pack at 4 °C Stability & Storage Keep as concentrated solution.

Store at 2~8°C and protected from prolonged exposure to light.Do not freeze.

This product is guaranteed up to one year from purchase.



Tel:240-252-7368(USA) Fax: 240-252-7376(USA) techsupport@elabscience.com Website: www.elabscience.com

Fluorophore

Conjugation: Elab Fluor® 488

Elab Fluor[®] 488 is designed to be excited by the Blue laser (488 nm) and detected using an optical filter centered near 520 nm (e.g., a 525/40 nm bandpass filter).

Recommended usage

Each lot of this antibody is quality control tested by flow cytometric analysis. The amount of the reagent is suggested to be used 5 μ L of antibody per test (million cells in 100 μ L staining volume or per 100 μ L of whole blood). Please check your vial before the experiment. Since applications vary, the appropriate dilutions must be determined for individual use.

Related Information

- 1. Sample Preparation for Flow Cytometry https://www.elabscience.com/List-detail-5594.html
- 2. Staining Cell Surface Targets for Flow Cytometry https://www.elabscience.com/List-detail-5568.html
- 3. Flow Cytometry Troubleshooting Tips https://www.elabscience.com/List-detail-5593.html
- 4. How to select the appropriate detection channel through the spectrogram? https://www.elabscience.com/List-detail-459742.html